

JA



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/642,756	08/19/2003	Shigenobu Sato	P24072	3012
7055	7590	02/24/2005	EXAMINER	
GREENBLUM & BERNSTEIN, P.L.C. 1950 ROLAND CLARKE PLACE RESTON, VA 20191			STEWART, ALVIN J	
			ART UNIT	PAPER NUMBER
			3738	
DATE MAILED: 02/24/2005				

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

10/642,756

Applicant(s)

SATO ET AL.

Examiner

Alvin J Stewart

Art Unit

3738

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 06 December 2004.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-6 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-6 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 19 August 2003 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☒ All b) ☐ Some * c) ☐ None of:
- 1) ☒ Certified copies of the priority documents have been received.
 - 2) ☐ Certified copies of the priority documents have been received in Application No. _____.
 - 3) ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date _____
- 5) ☐ Notice of Informal Patent Application (PTO-152)
- 6) ☐ Other: _____

Response to Arguments

Applicant's arguments with respect to claims 1-6 have been considered but are moot in view of the new ground(s) of rejection.

Claim Rejections - 35 USC § 112

The following is a quotation of the first paragraph of 35 U.S.C. 112:

The specification shall contain a written description of the invention, and of the manner and process of making and using it, in such full, clear, concise, and exact terms as to enable any person skilled in the art to which it pertains, or with which it is most nearly connected, to make and use the same and shall set forth the best mode contemplated by the inventor of carrying out his invention.

Claims 2 and 4 are rejected under 35 U.S.C. 112, first paragraph, as failing to comply with the written description requirement. The claim(s) contains subject matter which was not described in the specification in such a way as to reasonably convey to one skilled in the relevant art that the inventor(s), at the time the application was filed, had possession of the claimed invention. Nowhere in the specification discloses a width of the middle of the spacer that is greater than a width of at least one end of the spacer. Applicant must show support of this in the specification or delete the above structure limitations from claim 2.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claim 1 is rejected under 35 U.S.C. 102(b) as being anticipated by Lin US Patent 6,080,158.

Art Unit: 3738

Lin discloses an intervertebral implant (100) comprising a body defined by a pair of upper (120) and lower (130) surfaces and a pair of side surfaces (160) and a withdrawal preventer (see attachment) comprising a plurality of linear claw portions (121). Finally, the body has a distance between the upper and lower surfaces at the front side of the intervertebral spacer is greater than a distance between the upper and lower surfaces at the rear side of the intervertebral spacer (see Fig. 2).

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 2 and 4 are rejected under 35 U.S.C. 103(a) as being unpatentable over Lin US Patent 6,080,158 in view of Henry et al US Patent 5,766,252

Lin discloses the invention substantially as claimed. However, Lin does not disclose a width in the middle of the spacer that is greater than a width of at least one end of the spacer.

Henry et al teaches an implant comprising a width that is taper from the back of the implant to the front for the purpose of facilitating the insertion of the implant into the vertebral cavity (see Fig. 1).

It would have been obvious to one having ordinary skill in the art at the time the invention was to modify the implant of the Lin reference with the taper implant of the Henry et al reference in order facilitate the insertion of the implant into the vertebral cavity.

Regarding claim 4, see attachment showing a taper shape at the front side of the implant.

Claims 3 and 5 are rejected under 35 U.S.C. 103(a) as being unpatentable over Lin US Patent 6,080,158 in view of Lin US 6,325,827 B1.

Lin ('158) discloses the invention substantially as claimed. However, Lin does not disclose claws having asymmetric triangular shape with surfaces having different angles.

Lin ('827) teaches an implant comprising a plurality of asymmetrical claw portions having different angles for the purpose of facilitating the insertion of the implant and avoiding the expulsion of the implant when is already implanted (col. 5, lines 6-15).

It would have been obvious to one having ordinary skill in the art at the time the invention was to modify the claw portions of the Lin reference ('158) with the claw portions of the Lin reference ('827) in order to facilitate the insertion of the implant and avoid the expulsion of the implant when is already implanted.

Claim 6 is rejected under 35 U.S.C. 103(a) as being unpatentable over Lin US Patent 6,080,158 in view of Brantigan US Patent 5,425,772.

Lin discloses the invention substantially as claimed. However, Lin does not disclose a claw portion extending across the body from one side to the other.

Brantigan discloses an intervertebral implant comprising a plurality of claw portions extending across the body from one side to the other for the purpose of attaching the implant into adjoining vertebrae faces.

It would have been obvious to one having ordinary skill in the art at the time the invention was to modify the interrupted claw portions of the Lin reference with the non-interrupted claw portions of the Brantigan reference in order to attach the implant into adjoining vertebrae faces.

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Alvin J Stewart whose telephone number is 703-305-0277. The examiner can normally be reached on Monday-Friday 7:00AM-5:30PM(1 Friday B-week off).

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Corrine McDermott can be reached on 703-308-2111. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Art Unit: 3738

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

A handwritten signature in black ink, appearing to read "A. Stewart".

Alvin J Stewart
Primary Examiner
Art Unit 3738

February 19, 2005.